




SECTION THROUGH SUPPORT SYSTEM




(A)



(B)



(C)



(D)

NOTES:

ALL SUPPORTS AND HANGERS SHALL BE ARRANGED AS SHOWN ON ARCHITECTURAL DRAWINGS AND SHALL TERMINATE APPROX. 6" [150mm] FROM FACE OF WALLS OR PARTITIONS, EXCEPT FOR MEMBERS FLUSH WITH CEILING.

THE NUMBER SHOWN ON DRAWING WITH A "P" PREFIX ARE UNISTRUT CATALOG NUMBERS USED AS A CONVENIENCE FOR IDENTIFYING DESIRED SHAPES AND THICKNESS OF MEMBERS.

NOTE FOR ARCHITECT:

DETAILS A, B, C & D SHOW ANCHORAGE SUPPORT INTO VARIOUS TYPES OF SLAB CONDITIONS. ACTUAL SUPPORT MUST BE DESIGNED TO SUIT THE PARTICULAR SITUATION.

DETAIL TITLE / DOD / VA UNIVERSAL X-RAY (R-F) ROOM SUPPORT SYSTEM FOR CEILING MOUNTED X-RAY EQUIPMENT

Department of Veterans Affairs



SCALE : NONE

DATE ISSUED :DECEMBER 2008

CADD DETAIL NO. SD134900-06.DWG